

**NEO CALIBRATOR**  
**IMMUNOGLOBULIN-E (IgE)**  
**(Turbidimetry Method)**



Lot No. TCIgE0623A01 Mfg Dt – 06/2023  
EXP Dt – 11/2024

KIT NAME	KIT SIZE
NEO CAL - IgE	6 x 1 ml

**INTRODUCTION**

IgE Calibrator is intended to use as calibrator of accuracy and IgE for both semi and fully automated Chemistry Analyzer.

This IgE Calibrator is to be used in the same way as an unknown specimen that would be used in accordance with the direction of the IgE assay procedure.

PACKAGE 6 x 1 ml

C1 - IgE CALIBRATOR	1 x 1 ml
C2 - IgE CALIBRATOR	1 x 1 ml
C3 - IgE CALIBRATOR	1 x 1 ml
C4 - IgE CALIBRATOR	1 x 1 ml
C5 - IgE CALIBRATOR	1 x 1 ml
C6 - IgE CALIBRATOR	1 x 1 ml

Please refer the concentration on the vial label.

Product Name	Concentration
C1 - Level 1	0.0IU/ml
C2 - Level 2	50.0IU/ml
C3 - Level 3	100.0IU/ml
C4 - Level 4	200.0IU/ml
C5 - Level 5	500.0IU/ml
C6 - Level 6	1000.0IU/ml